


NANOCOLOR® Nitrat 8

Nitrate / Nitrato



Test 0-65
REF 985065

Methode / Method / Méthode / Método
0651 0.30–8.00 mg/L NO₃-N
0652 1.3–35.0 mg/L NO₃⁻

345 / 350 / 365 nm

Min	20–80 %		Max
0.30	1.84	6.46	8.00 mg/L NO ₃ -N
1.3	8.0	28.3	35.0 mg/L NO ₃ ⁻

