


NANOCOLOR[®] gesamt-Phosphat 45

total Phosphate / Phosphate total / Fosfato total













Test 0-55
REF 985055

Methode / Method / Méthode / Método

690 / 710 nm


0551 5.0–50.0 mg/L P
0552 15–150 mg/L PO₄³⁻
0553 10–115 mg/L P₂O₅

Min	20–80 %		Max
5.0	14.0	41.0	50.0 mg/L P
15	42	123	150 mg/L PO ₄ ³⁻



NANOCOLOR[®] ortho-Phosphat 45

ortho-Phosphate / ortho Phosphates / orto Fosfato



Test 0-55
REF 985055

Methode / Method / Méthode / Método

690 / 710 nm

0554 5.0–50.0 mg/L PO₄-P
0555 15–150 mg/L PO₄³⁻
0556 10–115 mg/L P₂O₅

Min	20–80 %		Max
5.0	14.0	41.0	50.0 mg/L PO ₄ -P
15	42	123	150 mg/L PO ₄ ³⁻

