

visicolor® Powder Pillows

Silica HR

Reagent set for the photometric determination of the silica content in water and sea water samples.

Measuring range:

2-210 mg/L SiO₂

1-100 mg/L Si

Method:

Photometric determination of the silica content using the silicomolybdate method analogous to APHA 4500-Si D.

Hazard warning:

Information regarding safety can be found on the box' label and in the safety data sheet. You can download the SDS from www.mn-net.com/SDS.

Procedure:

Requisite accessories: 2 test tubes 16 mm OD (REF 91680) or 2 test tubes 24 mm OD (REF 936101), special filter 450 nm

- 1 Rinse test tube several times with sample (*pH value of sample must be between pH 3 and 13*)

Blank (optional):

- 2 Fill one test tube with 5 mL of sample
- 3 Clean test tube
- 4 Place test tube in photometer as blank value and adjust for zero

Sample:

- 5 Fill another test tube with 5 mL of sample
- 6 Add content of 1 Powder Pillow "molybdate reagent"
- 7 Add content of 1 Powder Pillow "acid reagent"
- 8 Close test tube and vigorously shake until the solid material has dissolved
- 9 Wait for 3 min
- 10 Add content of 1 Powder Pillow "citric acid reagent"
- 11 Close test tube and shake well
- 12 Clean test tube
- 13 Wait for a 2 min reaction time
- 14 Measure

Measurement:

See manual for all MACHEREY-NAGEL photometers.

It should be ensured that any distilled water used for dilution is silica-free (REF 918912).

After use, rinse out test tubes thoroughly and seal them.

This method is also suitable for the analysis of sea water.

Interferences:


The following will not interfere: < 600 mg/L PO₄³⁻

The following will interfere: large amounts of Fe^{2+/3+}, oxidising agents, sulphides




According to APHA 4500-Si D, there is a modification of the silica which does not react with molybdate. This molybdate-unreactive form can be converted into the reactive species through heating or fusing with a base (e.g. digestion with sodium bicarbonate NaHCO₃).

Disposal of samples:

Information regarding disposal can be found in the safety data sheet. You can download the SDS from www.mn-net.com/SDS.



Blank (optional):



Sample:

