


visocolor® ECO Chlor 1 (frei) Chlorine (free)/Chlore (libre)/Cloro (libre)



Messbereich / Range / Domaine de mesure / Rango
0.1–2.0 mg/L Cl₂


REF 931 035
931 235




1
2 x 5 mL Probe
Sample
Echantillon
Muestra




2
1  Cl₂-1



3
20 s Schütteln
Shake
Agiter
Agitar



4
1'00 min



5
Messung
Measurement
Mesure
Medición




Meerwasser
Sea water
Eau de mer
Agua de mar


visocolor® ECO Chlor 1 (gesamt) Chlorine (total)/Chlore (totale)/Cloro (total)

Messbereich / Range / Domaine de mesure / Rango
0.1–2.0 mg/L Cl₂


REF 931 035
931 235




2 x 5 mL Probe
Sample
Echantillon
Muestra




1 Cl₂-1




20 s Schütteln
Shake
Agiter
Agitar




3 Cl₂-2



Schütteln
Shake
Agiter
Agitar



2'00 min



Messung
Measurement
Mesure
Medición



Meerwasser
Sea water
Eau de mer
Agua de mar