


NANOCOLOR[®] gesamt-Phosphat LR 1

total Phosphate / Phosphate total / Fosfato total













Test 0-95
REF 985095

Methode / Method / Méthode / Método

0951 0.05–0.50 mg/L P
0952 0.2–1.5 mg/L PO₄³⁻


690 / 885 nm

Min	20–80 %		Max
0.05	0.14	0.41	0.50 mg/L P
0.2	0.5	1.2	1.5 mg/L PO ₄ ³⁻



NANOCOLOR[®] ortho-Phosphat LR 1

ortho-Phosphate / ortho Phosphates / orto Fosfato



Test 0-95
REF 985095

Methode / Method / Méthode / Método

0955 0.05–0.50 mg/L PO₄-P

0956 0.2–1.5 mg/L PO₄³⁻

690 / 885 nm

Min	20–80 %		Max
0.05	0.14	0.41	0.50 mg/L PO ₄ -P
0.2	0.5	1.2	1.5 mg/L PO ₄ ³⁻

